



NOTES:

1. Connection for data communication to be made using conduit or watertight flex into ELECTRICAL INTERFACE ENCLOSURE, following diagram detail.
2. 15 amp minimum service provides power to the analyzer. Second convenience outlet required adjacent to analyzer as shown.
- 3.) All system components are less than 11.0" deep. Allow 31.0" minimum clearance for door swings.

ChemScan[®]
UV-2150/S Analyzer System
Typical Installation Diagram
- Front View, Wall Mount - Dual Inlet

ASA, Inc.	PHONE: 262-717-9500 FAX: 262-717-9530
Applied Spectrometry Associates, Inc. ©'08	
DATE: 1/12/07 REV: 7/30/08 BY: LJB	
FILE # 22SCWPFA	